

TRANSDUCER VERSATORR

BIBVT200

Brooks Transducer BVT200

- Type: Pirani Transducer
- Temperatur: Ambient -50°C
- Tryk område: 7.5x10⁻⁷ Torr igennem 1000 Torr
- Digital kommunikation: RS-232 /// RS-485



PRODUKTBESKRIVELSE

Software findes her. <https://www.brooksinstrument.com/en/products/accessories-software/product-software/versatorr-gui>

SPECIFIKATIONER

Analog output	0.5 - 9.5 Vdc, 1.0-9 Vdc, 0.375 - 5.659 Vdc, 1.9 - 10 Vdc, 1.5 - 8.5 Vdc. 1.9 - 9.1 Vdc, 0-10Vdc, 0-5Vdc
Certifikater	UL, CSA, EN/IEC, CE, RoHS
Digital output	RS-232, RS-485
Elektrisk tilslutning	9-Pin SUB-D Male, 15-Pin SUB-D Male, 15-Pin SUB-D Dual, 6-Pin Hirschmann, 8-Pin RJ45/FCC68
Funktion	Pirani Transducer
Medietemperatur fra	-20 °C
Medietemperatur til	50 °C
Nøjagtighed	0.25% of reading
Responstid	<20 ms
Temperaturområde fra	-20 °C
Temperaturområde til	50 °C
Tilslutning	DN16KF, DN25KF, 1/8" NPT, VCR4
Trykområde max	4 bar